

### Report on Outcome of Test of Pressure Durability

Application No. ..... Date/Month/Year as Tested .....

Name or Trademark..... Model .....

Meter as Tested No. ..... the Number of Water Meters as Tested.....Meters

The water meter of diameter less than 50 centimetres (shall be durable under the pressure at least 1,000 kPa.).

The water meter of diameter starting from 50 centimetres upward (shall be durable under the pressure at least 600 kPa.).

No.	Series Number of Instrument	Outcome of Test (Pressure Range for not less than 1 Minute)	
		<input type="radio"/> Pass	<input type="radio"/> not Pass
1.		<input type="radio"/>	<input type="radio"/>
2.		<input type="radio"/>	<input type="radio"/>
3.		<input type="radio"/>	<input type="radio"/>
4.		<input type="radio"/>	<input type="radio"/>
5.		<input type="radio"/>	<input type="radio"/>
6.		<input type="radio"/>	<input type="radio"/>
7.		<input type="radio"/>	<input type="radio"/>
8.		<input type="radio"/>	<input type="radio"/>
9.		<input type="radio"/>	<input type="radio"/>
10.		<input type="radio"/>	<input type="radio"/>
11.		<input type="radio"/>	<input type="radio"/>
12.		<input type="radio"/>	<input type="radio"/>
13.		<input type="radio"/>	<input type="radio"/>
14.		<input type="radio"/>	<input type="radio"/>
15.		<input type="radio"/>	<input type="radio"/>
16.		<input type="radio"/>	<input type="radio"/>
17.		<input type="radio"/>	<input type="radio"/>
18		<input type="radio"/>	<input type="radio"/>
19.		<input type="radio"/>	<input type="radio"/>
20.		<input type="radio"/>	<input type="radio"/>

**Remark :** In the case where the outcome of the test of the water meters as taken for the sampling is that some or all of the water meters do not pass the test, the test of all the water meters which have not been taken for the sampling shall be done.

To sum up the outcome of the test of pressure durability of the water meter :  Pass  not Pass

(Signed).....Licensee or Person Who Is Authorized

(.....)