

Unofficial Translation^{*}

Department of Commercial Registration

Notification

Re: Amendment of Weighing Instrument type 6

Since Weights and Measures Inspector found that metal plate on the top of center vertical spring weighing instrument type 6 if one of screw that hold tightly with body has loosen. It will be affect to the accuracy of instrument immediately and it will be clause from disloyal merchant to use this method apply for imprecise to exploit the populace. Department of commercial registration has purposed to amend and improve vertical spring weighing instrument type 6 (dial, mirror) to be more durable and accuracy on standard which protect instrument fraudulent by applied the screw on the top of center vertical spring weighing instrument must be hold tightly and cannot be loosen.

Therefore, if the owner of weighing instrument type 6 which the screw on the top of center vertical spring weighing instrument might be loosen following the officer of Central Bureau of Weights and Measures has surveyed on 29 July Buddhist Era 2518 which has not apply for certificate yet, must be apply for occupied instrument as soon as possible. Inspector of Central Bureau of Weights and Measures will inspect and certify until it complete and weighing instrument which apply for new certificate must follow the regulation.

Declared to be informed accordingly

Given on the 8 Day of August B.E. 2519

(Signature) Chai Nitiphrapa

(Mr. Chai Nitiphrapa)

Director General

Department of Commercial Registration